


# Mc-THERMO VIN Kabel offen Mc-THERMO VIN Cable open




Bestell-Nr.  620-625.005 (0,5m)  620-625.050 (5,0m)  
Order-N<sup>o</sup>  620-625.015 (1,5m)  620-625.100 (10,0m)  
 620-625.030 (3,0m)  620-625.

## 620-625.



THERMO-SUB-MIN "rot"  
Mc-THERMO



## Sicherheitshinweis:

Die offenen Kontakte des ausgesteckten Thermosteckers können einen Kurzschluss des spannungsführenden Messpunktes verursachen. Hierdurch können erhebliche Schäden am Kabel und dem Messobjekt entstehen.

Trennen Sie das Kabel von der Messstelle, bevor Sie das Messkabel am Modul abziehen!

## Safety Warning:

The blank lugs of the unplugged thermo connector could short-circuit the voltage of the measuring point of which the cable is connected to.

This might damage the cable and the measuring object as well.

Disconnect the cable from the measuring point before you pull off the thermo connector from the module!

620-625\_by\_bpz.prt 10.05.10 IH

Alle Rechte vorbehalten. Nachdruck und Vervielfältigung, auch auszugsweise sind verboten!  
All rights reserved. Reproduction or duplication in part or in whole are forbidden.